



Designation: D1385 – 07 (Reapproved 2018)^{ε1}

Standard Test Method for Hydrazine in Water¹

This standard is issued under the fixed designation D1385; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

^{ε1} NOTE—Warning notes were editorially updated throughout and the Keywords section was added in August 2018.

1. Scope

1.1 This test method covers² the colorimetric determination of hydrazine in boiler feed waters, condensates, natural, and well waters that have been treated with hydrazine (N₂H₄). This test method is usable in the range from 5.0 to 200 µg/L (ppb) hydrazine. The range is for photometric measurements made at 458 nm in 50 mm cell. Higher concentrations of hydrazine can also be determined by taking a more diluted sample.

1.2 It is the users' responsibility to ensure the validity of this test method for untested types of waters.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific precautionary statements, see 5.3, 8.4, and Footnote 5.

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee D19 on Water and is the responsibility of Subcommittee D19.03 on Sampling Water and Water-Formed Deposits, Analysis of Water for Power Generation and Process Use, On-Line Water Analysis, and Surveillance of Water

Current edition approved Aug. 1, 2018. Published September 2018. Originally approved in 1967. Last previous edition approved in 2013 as D1385 – 07 (2013)ε1. DOI: 10.1520/D1385-07R18E01.

² For further information on this test method, the following references may be of interest:

Watt, G. W., and Chrisp, J. D., "Spectrophotometric Method for the Determination of Hydrazine," *Analytical Chemistry*, Vol 24, No. 12, 1952, pp. 2006–2008; and Wood, P. R., "Determination of Maleic Hydrazide Residues in Plant and Animal Tissue," *Analytical Chemistry*, Vol 25, No. 12, 1953, pp. 1879–1883.

2. Referenced Documents

2.1 ASTM Standards:³

D1066 Practice for Sampling Steam
D1129 Terminology Relating to Water
D1193 Specification for Reagent Water
D3370 Practices for Sampling Water from Closed Conduits
D5810 Guide for Spiking into Aqueous Samples
D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis
E60 Practice for Analysis of Metals, Ores, and Related Materials by Spectrophotometry
E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this standard, refer to Terminology D1129.

4. Summary of Test Method

4.1 When a solution of p-dimethylaminobenzaldehyde in methyl alcohol and hydrochloric acid is added to hydrazine in diluted hydrochloric acid solution, a characteristic yellow color of p-dimethylaminobenzalazine is formed. The yellow color formed is proportional to the hydrazine present and is in good agreement with Beer's law in the range from 5.0 to 200 µg/L (ppb) hydrazine.

5. Significance and Use

5.1 Hydrazine is a man-made chemical and is not found in natural waters. The determination of hydrazine is usually made on boiler feedwaters, process waters, and other waters that have been treated with hydrazine (N₂H₄) for the purpose of maintaining residuals to prevent corrosion by dissolved oxygen. This reducing chemical reacts with dissolved oxygen to form nitrogen and water. However, under certain conditions it

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

can also decompose to form ammonia and nitrogen. Hydrazine is used extensively as a preboiler treatment chemical for high-pressure boilers to scavenge small amounts of dissolved oxygen that are not removed by mechanical aeration. It has the advantage over sulfite treatment in that it does not produce any dissolved solids in the boiler water. Hydrazine is often determined in concentrations below 0.1 mg/L. However, in layup solutions for the protection of idle boilers, hydrazine may be present in concentrations as high as 200 mg/L.

5.2 Additionally, hydrazine provides protection where reducing conditions are required, particularly in mixed metal-lurgy systems for the protection of the copper alloys.

5.3 Hydrazine is a suspected carcinogen and a threshold limit value in the atmosphere of 1.0 mg/L has been set by OSHA. When in an aqueous solution, hydrazine will oxidize to nitrogen and water in the presence of air over a relatively short period of time.

6. Interferences

6.1 The substances normally present in industrial water do not interfere with the test; however, the hydrazine content may be diminished by oxidizing agents, such as chlorine, bromine, and iodine, collected with the sample or absorbed by it prior to testing.

6.2 Colors in the prescribed wavelengths also interfere, as do other dark colors or turbidities that cannot be overcome.

6.3 Aromatic amines, such as aniline, will also interfere.

7. Apparatus

7.1 *Photometer*—A spectrophotometer suitable for measurements at 458 nm and capable of holding cells with a light path of 50 mm should be used. Filter photometers and photometric practices prescribed in this test method shall conform to Practice E60, and spectrophotometers to Practice E275.

7.2 Certain photoelectric filter photometers are capable of measurement at 425 nm, but not at 458 nm. Measurements may be made at 425 nm with a reduction in sensitivity of approximately 50 % of that possible at 458 nm.

7.3 Instruments that read out in direct concentration can also be used. Manufacturer's instructions should be followed.

8. Reagents

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁴ Other grades may be used, provided it is first ascertained that the reagent is sufficiently high in purity to permit its use without lessening the accuracy of the determinations.

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to the quantitative requirements of Type III reagent water in Specification D1193.

8.3 *Hydrazine Solution, Stock* (1.0 mL = 100 µg N₂H₄)—Dissolve 0.328 g of hydrazine dihydrochloride (HCl·NH₂·NH₂·HCl) in 100 mL of water and 10 mL of HCl (sp gr 1.19). Dilute with water to 1 L in a volumetric flask and mix. (**Warning**—See 8.4.)

8.4 *Hydrazine Solution, Standard* (1.0 mL = 0.500 µg N₂H₄)—Dilute 5.0 mL of hydrazine stock solution to 1 L with water and mix. Prepare as needed. (**Warning**—Hydrazine is a suspected carcinogen and should be handled with care.⁵)

8.5 *Hydrochloric Acid* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

8.6 *p-Dimethylaminobenzaldehyde Solution*—Dissolve 4.0 g of p-dimethylaminobenzaldehyde [(CH₃)₂NC₆H₄CHO] in 200 mL of methyl alcohol (CH₃OH) and 20 mL of HCl (sp gr 1.19). Store in a dark bottle out of direct sunlight.

9. Sampling

9.1 Collect the sample in accordance with Practices D3370 or Practice D1066, whichever is applicable. (**Warning**—See 8.4.)

9.2 Acidify and dilute the sample as soon as taken by adding 1 mL of concentrated HCl (sp gr 1.19) to a 100-mL volumetric flask and then pipetting 50 mL of the sample into the flask and diluting to 100 mL. Prepare a blank with water at the same time.

9.3 A smaller sample aliquot should be taken if the hydrazine concentration is greater than 200 µg/L.

10. Calibration

10.1 Prepare a series of standard hydrazine solutions by pipetting 0.0, 5.0, 10.0, 25.0, 50.0, 100.0, and 200.0 mL of hydrazine standard solution (1.0 mL = 0.500 µg N₂H₄) into 500-mL volumetric flasks. Add 5 mL of HCl (sp gr 1.19) to each flask and dilute with water to 500 mL and mix well. This will give standard solutions containing 0, 5.0, 10.0, 25.0, 50.0, 100, and 200 µg/L (ppb) of hydrazine.

10.2 Pipet 50.0-mL portions of the hydrazine standard solutions into clean, dry 100-mL beakers or flasks and proceed as directed in 11.2. Plot absorbance on the ordinate and micrograms per litre of hydrazine on the abscissa of linear graph paper. Alternately, graph the data in an electronic spreadsheet or use an instrument that reads out in direct concentrations.

10.3 A separate calibration curve must be made for each photometer and a recalibration must be made if it is necessary to change the cell, lamp, or filter, or if any other alterations of instrument or reagents are made. Check the curve for each

⁵ MacEwen, J. D., Vernot, E. H., Haun, C. C., and Kinkead, E. B., "Chronic Inhalation Toxicity of Hydrazine: Onconogenic Effects," in cooperation with the University of California (Irvine) and the Airforce Aero Medical Research Laboratory.