



Designation: D5845 – 21

Standard Test Method for Determination of MTBE, ETBE, TAME, DIPE, Methanol, Ethanol and *tert*-Butanol in Gasoline by Infrared Spectroscopy¹

This standard is issued under the fixed designation D5845; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination of methanol, ethanol, *tert*-butanol, methyl *tert*-butyl ether (MTBE), ethyl *tert*-butyl ether (ETBE), *tert*-amyl methyl ether (TAME), and diisopropyl ether (DIPE) in gasoline by infrared spectroscopy. The test method is suitable for determining methanol from 1.7 % to 5.5 % by mass, ethanol from 1.5 % to 10 % by mass, *tert*-butanol from 1.6 % to 12 % by mass, DIPE from 1.2 % to 17 % by mass, MTBE from 2.0 % to 16.4 % by mass, ETBE from 2.0 % to 18.5 % by mass, and TAME from 1.5 % to 18.5 % by mass.

1.2 This test method is applicable to oxygenate(s) in gasoline singularly or in multiple mixtures hereof according to the oxygenates and mass percent ranges given in 1.1.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- D1298 Test Method for Density, Relative Density, or API Gravity of Crude Petroleum and Liquid Petroleum Products by Hydrometer Method
- D4052 Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter
- D4057 Practice for Manual Sampling of Petroleum and Petroleum Products
- D4307 Practice for Preparation of Liquid Blends for Use as Analytical Standards
- D4815 Test Method for Determination of MTBE, ETBE, TAME, DIPE, tertiary-Amyl Alcohol and C₁ to C₄ Alcohols in Gasoline by Gas Chromatography
- D5599 Test Method for Determination of Oxygenates in Gasoline by Gas Chromatography and Oxygen Selective Flame Ionization Detection
- D8321 Practice for Development and Validation of Multivariate Analyses for Use in Predicting Properties of Petroleum Products, Liquid Fuels, and Lubricants based on Spectroscopic Measurements

NOTE 1—D5845 was developed with reference to E1655. A significant enhancement has been made to E1655 and is designated as D8321. D8321 is the preferred standard Practice and describes the mathematics, that is, the regression algorithms to develop calibration equations that are also described in E1655. Replace E1655 with D8321.

NOTE 2—Surrogate methods such as D5845 may indicate that they make use of the creation of multivariate calibration equations described therein D8321, but they should not claim to follow the procedures described herein.

2.2 Other Standard:³

GC/OFID EPA Test Method—Oxygen and Oxygenate Content Analysis (by way of gas chromatography with oxygen-selective flame ionization detection)

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.04.0F on Absorption Spectroscopic Methods.

Current edition approved Oct. 1, 2021. Published November 2021. Originally approved in 1995. Last previous edition approved in 2016 as D5845 – 01 (2016). DOI: 10.1520/D5845-21.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Code of Federal Regulations, Part 80 of Title 40, Section 80.46(g); also published in the Federal Register, Volume 59, No. 32, February 16, 1994, p 7828.

*A Summary of Changes section appears at the end of this standard

3. Terminology

3.1 Definitions:

3.1.1 *diluent, n*—a substance used to dilute something.

3.1.1.1 *Discussion*—In this standard, the diluents are termed base gasoline A, B, and C, which are the diluents consisting of gasoline blend stocks used to dissolve the oxygenates to create the calibration and instrument qualification matrix solutions. The base stocks must not contain measurable oxygenates. The known base socks are:

(1) Base gasoline A should be a gasoline with at least 60 % alkylate. A suggested recipe for base gasoline A is 60 % alkylate, 30 % full range reformat, and 10 % light straight run.

(2) Base gasoline B should be a gasoline with at least 60 % full range reformat. A suggested recipe for base gasoline B is 60 % full range reformat, 30 % FCC gasoline, and 10 % light straight run.

(3) Base gasoline C the composition is not known

3.1.2 *multivariate calibration, n*—a process for creating a calibration model in which multivariate mathematics is applied to correlate the absorbances measured for a set of calibration samples to reference component concentrations or property values for the set of samples. The resultant multivariate calibration model is applied to the analysis of spectra of unknown samples to provide an estimate of the component concentration or property values for the unknown sample.

3.1.3 *oxygenate, n*—a molecule composed solely of carbon, hydrogen, and oxygen.

3.1.3.1 *Discussion*—In this standard, the oxygenates of interest are: methanol, ethanol, *tert*-butanol, DIPE, MTBE, ETBE, and TAME.

4. Summary of Test Method

4.1 A sample of gasoline is introduced into a liquid sample cell. A beam of infrared light is imaged through the sample onto a detector, and the detector response is determined. Regions of the infrared spectrum are selected for use in the analysis by either placing highly selective bandpass filters before or after the sample or mathematically selecting the regions after the whole spectrum is obtained. A multivariate mathematical analysis is carried out which uses the detector response for the selected regions in the spectrum to compute the concentration for each component in an unknown based on multivariate models for each component.

5. Significance and Use

5.1 Alcohols and ethers are added to gasoline to produce a reformulated lower emissions gasoline. Alcohols and ethers may also be added to gasoline to increase the octane number. Type and concentration of various oxygenates are specified and regulated to ensure acceptable commercial gasoline quality. Driveability, vapor pressure, phase separation, and evaporative emissions are some of the concerns associated with oxygenated fuels.

5.2 This test method is faster, simpler, less expensive and more portable than current methods.

5.3 This test method may be applicable for quality control in the production of gasoline.

5.4 This test method is not suitable for testing for compliance with federal regulations.³

5.5 False positive readings for some of the samples tested in the round robin were sometimes observed. As only extreme base gasolines were tested in the round robin, no definitive statement can be made as to the expected frequency or magnitude of false positives expected in a wider range of base gasolines.

6. Apparatus

6.1 *Mid-IR Spectrometric Analyzer*, of one of the following types:

6.1.1 *Filter-based Mid-IR Test Apparatus*—The type of apparatus suitable for use in this test method minimally employs an IR source, an infrared transmission cell or a liquid attenuated total internal reflection cell, wavelength discriminating filters, a chopper wheel, a detector, an A-D converter, a microprocessor, and a sample introduction system.

6.1.2 *Fourier Transform Mid-IR Test Apparatus*—The type of apparatus suitable for use in this test method employs an IR source, an infrared transmission cell or a liquid attenuated total internal reflection cell, a scanning interferometer, a detector, an A-D converter, a microprocessor and a sample introduction system.

6.1.3 *Dispersive Mid-IR Test Apparatus*—The type of apparatus suitable for use in this test method minimally employs an IR source, an infrared transmission cell or a liquid attenuated total internal reflection cell, a wavelength dispersive element such as a grating or prism, a chopper wheel, a detector, an A-D converter, a microprocessor and a sample introduction system.

7. Reagents and Materials

7.1 *Samples for Calibration and Quality Control Check Solutions*—Use of chemicals of at least 99 % purity is highly recommended when preparing calibration and quality control check samples. If reagents of high purity are not available, an accurate assay of the reagent must be performed using a properly calibrated GC or other techniques (for example, water determination).

7.1.1 Base gasolines containing no oxygenates,

7.1.2 Methanol,

7.1.3 Ethanol,

7.1.4 *tert*-Butanol,

7.1.5 Methyl *tert*-butyl ether, MTBE,

7.1.6 Ethyl *tert*-butyl ether, ETBE,

7.1.7 *tert*-Amyl methyl ether, TAME, and

7.1.8 Diisopropyl ether, DIPE.

7.2 **Warning**—These materials are flammable and may be harmful if ingested or inhaled.

8. Sampling and Sample Handling

8.1 General Requirements:

8.1.1 Gasoline samples must be handled with meticulous care to prevent evaporative loss and composition changes.

8.1.2 Gasoline samples to be analyzed by the test method shall be obtained using method(s) specified by governmental regulatory agencies or by the procedures outlined in Practice **D4057** (or equivalent). Do not use the “Sampling by Water