



Designation: D5932 – 20

Standard Test Method for Determination of 2,4-Toluene Diisocyanate (2,4-TDI) and 2,6-Toluene Diisocyanate (2,6-TDI) in Air (with 9-(N-Methylaminomethyl) Anthracene Method) (MAMA) in the Workplace¹

This standard is issued under the fixed designation D5932; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of gaseous 2,4-toluene diisocyanate (2,4-TDI) and 2,6-toluene diisocyanate (2,6-TDI) in air samples collected from workplace and ambient atmospheres.

1.2 Differential air sampling is performed with a segregating device.² The gaseous fraction is collected on a glass fiber filter (GFF) impregnated with 9-(N-methylaminomethyl) anthracene (MAMA).

1.3 The analysis of the gaseous fraction is performed with a high performance liquid chromatograph (HPLC) equipped with ultraviolet (UV) and fluorescence detectors. An ultra high performance liquid chromatograph (UPLC) can also be used, provided that its performance is equivalent to what is stated in this standard.

1.4 The analysis of the aerosol fraction is performed separately as described in Ref (1).³

1.5 The range of application of this test method, utilizing UV and a fluorescence detector, is validated for 0.014 to 1.16 μg of monomer 2,4- and 2,6-TDI/2.0 mL of desorption solution, which corresponds to concentrations of 0.001 to 0.077 mg/m^3 of TDI based on a 15-L air sample. This corresponds to 0.014 to 11 ppb(V) and brackets the established TLV value of 1 ppb(v).

1.6 A field blank sampling system is used to check the possibility of contamination during the entire sampling and analysis.

1.7 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use. See Section 9 for additional hazards.*

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:⁴

- D1193 Specification for Reagent Water
- D1356 Terminology Relating to Sampling and Analysis of Atmospheres
- D1357 Practice for Planning the Sampling of the Ambient Atmosphere
- D4840 Guide for Sample Chain-of-Custody Procedures
- D5337 Practice for Flow Rate Adjustment of Personal Sampling Pumps

2.2 Other Documents:

- Sampling Guide for Air Contaminants in the Workplace⁵

3. Terminology

3.1 For definitions of terms used in this test method, refer to Terminology D1356.

¹ This test method is under the jurisdiction of ASTM Committee D22 on Air Quality and is the direct responsibility of Subcommittee D22.04 on Workplace Air Quality.

Current edition approved March 1, 2020. Published May 2020. Originally approved in 1996. Last previous edition approved in 2013 as D5932 – 08 (2013) ϵ^1 . DOI: 10.1520/D5932-20.

² ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

³ The boldface numbers in parentheses refer to a list of references at the end of this standard.

⁴ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

⁵ Available from Institut de Recherche en Santé et en Sécurité du Travail (IRSST), Laboratory Division, Montreal, QC, 2012, <http://www.irsst.qc.ca/media/documents/Pubirsst/T-15.pdf>.


4. Summary of Test Method

4.1 A known volume of air is drawn through a segregating sampling device.

4.2 Gaseous and aerosol fraction are sampled simultaneously with a two filter loaded cassette. The aerosol is collected on the first filter made of polytetrafluoroethylene (PTFE), the gaseous counterpart being adsorbed on the second filter made of glass fiber (GFF) impregnated with MAMA.

4.3 The analysis of the monomer and oligomer in the aerosol fraction is performed separately in accordance with the procedure described in Refs (1, 2).

4.4 The diisocyanate present as a gas reacts with the secondary amine function of the MAMA impregnated on the GFF to form a urea derivative (3, 4), as shown below.



4.5 Desorption is done with dimethylformamide 67 % containing 33 % mobile phase (70 % acetonitrile, 30 % buffer).

4.6 The resulting solution is analyzed by HPLC with two detectors in series: UV (254 nm) and fluorescence (254-nm excitation and 412-nm emission) (5).

4.7 2,4- and 2,6-TDI urea derivatives are separated using a reversed phase HPLC column.

4.8 A complete calibration curve, covering the range of application of the test method, was obtained to determine the linearity of the method (see 1.5).

4.9 Concentration of urea derivative contained in the samples is calculated by using an external standard of the appropriate urea derivative.

5. Significance and Use

5.1 TDI is used mostly in the preparation of rigid and semi-rigid foams and adhesives.

5.2 Diisocyanates and polyisocyanates are irritants to skin, eyes, and mucous membranes. They are recognized to cause respiratory allergic sensitization, asthmatic bronchitis, and acute respiratory intoxication (6-9).

5.3 The American Conference of Governmental Industrial Hygienists (ACGIH) has adopted a Threshold Limit Value-Time Weighted Average (TLV-TWA) of 0.001 ppm of 0.007 mg/m³ with a Short-Term Exposure Limit (STEL) of 0.005 ppm or 0.036 mg/m³ for either 2,4-TDI, or 2,6-TDI, or for a mixture of 2,4- and 2,6-TDI (10). The Occupational Safety and Health Administration of the U.S. Department of Labor (OSHA) has a permissible exposure limit of 0.02 ppm(V) or 0.14 mg/m³ of 2,4-TDI as a ceiling limit and 0.005 ppm (V) or 0.036 mg/m³ as a time-weighted average (11).

5.4 Monitoring of respiratory and other problems related to diisocyanates and polyisocyanates is aided through the utilization of this test method, due to its sensitivity and low volume requirements (15 L). Its short sampling times are compatible

with the duration of many industrial processes and its low quantification limit also suits the concentrations often found in the working area.

5.5 The segregating sampling device pertaining to this proposed test method physically separates gas and aerosol allowing isocyanate concentrations in both physical states to be obtained, thus helping in the selection of ventilation systems and personal protection.

5.6 This test method is used to measure gaseous concentrations of 2,4- and 2,6-TDI in air for workplace and ambient atmospheres.

6. Interference

6.1 Any substance that can react with MAMA reagent impregnated on the GFF can affect the sampling efficiency. This includes strong oxidizing agents.

6.2 Any compound that has the same retention time as the TDIU derivative and gives the same UV/fluorescence detector response factor ratio can cause interference. Chromatographic conditions can be changed to eliminate an interference.

6.3 A field blank double-filter sampling system is used to check contamination during the combined sampling, transportation, and sample storage process. A laboratory blank is used to check contamination occurring during the analytical process.

7. Apparatus

7.1 Sampling Equipment:

7.1.1 *Personal Sampling Pump*, capable of sampling 1.0 L/min or less for 4 h.

7.1.2 *Double Filter Sampling Device*, 37 mm in diameter, three-piece personal monitor, plastic holder loaded with a PTFE filter close to the mouth, followed by a glass fiber filter impregnated with MAMA and a plastic back-up pad. The glass fiber filter is impregnated with an amount of MAMA in the range of 0.07 to 0.25 mg.

7.1.3 Flow Measuring Device.

7.2 Analytical Equipment:

7.2.1 *Liquid Chromatograph*, a high-performance liquid chromatograph equipped with UV (254-nm wavelength) and fluorescence detectors (412-nm emission and 254-nm excitation) and an automatic or manual sample injector. A second UV wavelength is recommended for identify confirmation if no fluorescence detector is utilized. An ultra high performance liquid chromatography (UPLC) providing at least the same or equivalent performance of HPLC can be also used.

7.2.2 *Liquid Chromatographic Column*, an HPLC stainless steel column, capable of separating the urea derivatives. This proposed method recommends a 150- by 4.6-mm internal diameter stainless steel column packed with 3.5-μm C18, or an equivalent column.

7.2.3 *Electronic Integrator*, an electronic integrator or any other effective method for determining peak areas.

7.2.4 *Analytical Balance*, an analytical balance capable of weighing to 0.001 g.