



Designation: D6277 – 07 (Reapproved 2022)

Standard Test Method for Determination of Benzene in Spark-Ignition Engine Fuels Using Mid Infrared Spectroscopy¹

This standard is issued under the fixed designation D6277; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of the percentage of benzene in spark-ignition engine fuels. It is applicable to concentrations from 0.1 % to 5 % by volume.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- D1298 Test Method for Density, Relative Density, or API Gravity of Crude Petroleum and Liquid Petroleum Products by Hydrometer Method
- D4052 Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter
- D4057 Practice for Manual Sampling of Petroleum and Petroleum Products
- D4177 Practice for Automatic Sampling of Petroleum and Petroleum Products
- D4307 Practice for Preparation of Liquid Blends for Use as Analytical Standards

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.04.0F on Absorption Spectroscopic Methods.

Current edition approved May 1, 2022. Published May 2022. Originally approved in 1998. Last previous edition approved in 2017 as D6277 – 07 (2017). DOI: 10.1520/D6277-07R22.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- D5769 Test Method for Determination of Benzene, Toluene, and Total Aromatics in Finished Gasolines by Gas Chromatography/Mass Spectrometry
- D5842 Practice for Sampling and Handling of Fuels for Volatility Measurement
- D5854 Practice for Mixing and Handling of Liquid Samples of Petroleum and Petroleum Products
- E168 Practices for General Techniques of Infrared Quantitative Analysis
- E1655 Practices for Infrared Multivariate Quantitative Analysis
- E2056 Practice for Qualifying Spectrometers and Spectrophotometers for Use in Multivariate Analyses, Calibrated Using Surrogate Mixtures

3. Terminology

3.1 Definitions:

3.1.1 *multivariate calibration, n*—a process for creating a calibration model in which multivariate mathematics is applied to correlate the absorbances measured for a set of calibration samples to reference component concentrations or property values for the set of samples.

3.1.1.1 *Discussion*—The resultant multivariate calibration model is applied to the analysis of spectra of unknown samples to provide an estimate of the component concentration or property values for the unknown sample.

3.1.1.2 *Discussion*—Included in the multivariate calibration algorithms are Partial Least Squares, Multilinear Regression, and Classical Least Squares Peak Fitting.

3.1.2 *oxygenate, n*—an oxygen-containing organic compound which may be used as a fuel or fuel supplement, for example, various alcohols and ethers.

4. Summary of Test Method

4.1 A sample of spark-ignition engine fuel is introduced into a liquid sample cell. A beam of infrared light is imaged through the sample onto a detector, and the detector response is determined. Wavelengths of the spectrum, that correlate highly with benzene or interferences, are selected for analysis using selective bandpass filters or by mathematically selecting areas of the whole spectrum. A multivariate mathematical analysis converts the detector response for the selected areas of the spectrum of an unknown to a concentration of benzene.

5. Significance and Use

5.1 Benzene is a compound that endangers health, and the concentration is limited by environmental protection agencies to produce a less toxic gasoline.

5.2 This test method is fast, simple to run, and inexpensive.

5.3 This test method is applicable for quality control in the production and distribution of spark-ignition engine fuels.

6. Interferences

6.1 The primary spectral interferences are toluene and other monosubstituted aromatics. In addition, oxygenates can interfere with measurements made with filter apparatus. Proper choice of the apparatus, proper design of a calibration matrix, and proper utilization of multivariate calibration techniques can minimize these interferences.

7. Apparatus

7.1 *Mid-IR Spectrometric Analyzer (of one of the following types):*

7.1.1 *Filter-based Mid-IR Test Apparatus*—The type of apparatus suitable for use in this test method minimally employs an IR source, an infrared transmission cell or a liquid attenuated total internal reflection cell, wavelength discriminating filters, a chopper wheel, a detector, an A-D converter, a microprocessor, and a method to introduce the sample. The frequencies and bandwidths of the filters are specified in Table 1.

7.1.2 *Fourier Transform Mid-IR Spectrometer*—The type of apparatus suitable for use in this test method employs an IR source, an infrared transmission cell or a liquid attenuated total internal reflection cell, a scanning interferometer, a detector, an A-D converter, a microprocessor, and a method to introduce the sample. The following performance specifications (through the ATR cell) must be met:

scan range	4000 cm ⁻¹ to 600 cm ⁻¹
resolution	4 cm ⁻¹
S/N at 674 cm ⁻¹	>300:1 RMS

The signal to noise level will be established by taking a single beam spectrum using air or nitrogen as the reference and declaring that spectrum as the background. The background single beam spectrum obtained can be the average of multiple FTIR scans, but the total collection time shall not exceed 60 s. If interference from water vapor or carbon dioxide is a problem, the instrument shall be purged with dry air or nitrogen. A subsequent single beam spectrum shall be taken under the same conditions and ratioed to the background spectrum. The RMS noise of the ratioed spectra, the 100 %

line, shall not exceed 0.3 % transmittance in the region from 700 cm⁻¹ to 664 cm⁻¹.

7.2 *Absorption Cell*—The absorption cell can be either transmission or attenuated total reflectance.

7.2.1 *Transmission Cells*, shall have windows of potassium bromide, zinc selenide, or other material having a significant transmission from 712 cm⁻¹ to 660 cm⁻¹. The cell path length of the transmission cell shall be 0.025 mm ± 0.005 mm. The use of a wedged transmission cell with the same nominal path length is acceptable.

7.2.2 *Attenuated Total Reflectance (ATR) Cells*, shall have the following specifications:

ATR element material	ZnSe
beam condensing optics	conical, non-focussing optics integral to cell body
element configuration	circular cross section with coaxial conical ends
cone half angle	60°
element length	1.55 in.
element diameter	0.125 in.
angle of incidence at sample interface	53.8°
maximum range of incidence angles	± 1.5°
standard absorbance (1428 cm ⁻¹ band of acetone)	0.38 AU ± 0.02 AU
material of construction	316 stainless steel
seals	Chemraz or Kalraz o-rings

8. Reagents and Materials (see Note 1)

8.1 *Standards for Calibration, Qualification, and Quality Control Check Standards*—Use of chemicals of at least 99 % purity, where available, for quality control checks is required when preparing samples. (**Warning**—These materials are flammable and may be harmful if ingested or inhaled.)

8.1.1 *tert*-Amyl methyl ether, TAME [994-05-8].

8.1.2 Benzene [1076-43-3].

8.1.3 *tert*-Butyl ethyl ether, ETBE [637-92-3].

8.1.4 *tert*-Butyl methyl ether, MTBE [1634-04-4].

8.1.5 1,3 Dimethylbenzene (*m*-xylene).

8.1.6 Ethanol [64-17-5].

8.1.7 Ethylbenzene [100-41-4].

8.1.8 3-Ethyltoluene [620-14-4].

8.1.9 Heavy aromatic/reformate petroleum stream (high boiling cut: IPB of 150 °C ± 5 °C and EP of 245 °C ± 8 °C) certified to contain less than 0.025 % benzene (an absorbance of less than 0.03 at 675 cm⁻¹ using a 0.2 mm cell and a baseline between approximately 680 cm⁻¹ and 670 cm⁻¹) [64741-68-0].

8.1.10 Hexane (an absorbance versus water of less than 0.1 at 250 nm using a 1 cm cell) [110-54-3].

8.1.11 2,2,4-Trimethylpentane (*isooctane*) [540-84-1].

8.1.12 Pentane (an absorbance versus water of less than 0.1 at 250 nm using a 1 cm cell) [109-66-0].

8.1.13 Propylbenzene [103-65-1].

8.1.14 Toluene [108-88-3].

8.1.15 1,3,5-Trimethylbenzene (*mesitylene*) [108-67-8].

8.1.16 *m*-Xylene [108-38-3].

NOTE 1—Only some of the reagents are required in each calibration or qualification procedure.

9. Sampling and Sample Handling

9.1 General Requirements:

TABLE 1 Specification for Filters Used in Filter-based Mid-IR Test

Center Wavenumber (± 0.15 % of wavenumber)	Bandwidth (in wavelength units) (full width at half height)
673 cm ⁻¹	1 % of λ _c
729 cm ⁻¹	1 % of λ _c
769 cm ⁻¹	1 % of λ _c
1205 cm ⁻¹	1 % of λ _c
1054 cm ⁻¹	1 % of λ _c
1188 cm ⁻¹	1 % of λ _c
1117 cm ⁻¹	1 % of λ _c