



Designation: E1292 – 94 (Reapproved 2022)

Standard Specification for Gravity Convection and Forced Ventilation Incubators¹

This standard is issued under the fixed designation E1292; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers the performance requirements for general purpose air incubators ordinarily used for incubating procedures, which have an incubating chamber up to 0.6 m³ (25 ft³) in volume. It is applicable to gravity and forced ventilation incubators designed to operate over all or part of the temperature range from 5 °C above ambient to 75 °C.

1.2 This specification does not include any requirements for the safe handling of harmful or disease bearing organisms.

1.3 The following precautionary caveat pertains only to the test method portions, Sections 4, 5, and 6, of this specification. *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Types

2.1 This specification covers the following four types of air incubators:

2.1.1 *Type 1A*—An incubator ventilated by gravity convection having a uniformity of temperature within 3 % of the differential between incubator and ambient temperature,

2.1.2 *Type 1B*—An incubator ventilated by gravity convection having a uniformity of temperature within 6 % of the differential between incubator and ambient temperature,

2.1.3 *Type 2A*—An incubator ventilated by forced ventilation having a uniformity of temperature within 2 % of the differential between incubator and ambient temperature,

2.1.4 *Type 2B*—An incubator ventilated by forced ventilation having a uniformity of temperature within 5 % of the differential between incubator and ambient temperature.

3. Performance Requirements

3.1 The temperature within the incubating chamber shall be controlled by an automatic device, and shall be uniform within the tolerances given in Table 1, when tested in accordance with Section 4.

3.2 The rate of ventilation of the incubating chamber shall conform to the requirements specified in Table 1 for the particular type of incubator when measured in accordance with the procedure given in Section 5.

3.3 The recovery time of the incubating chamber shall conform to the requirements specified in Table 1 for the particular type of incubator, when tested in accordance with the procedure specified in Section 6.

4. Test Methods

4.1 Temperature Uniformity:

4.1.1 Place nine calibrated thermocouples (Note 1) made from iron-constantan or copper-constantan wires approximately 0.5 mm in diameter (No. 24 gauge) and having a junction size of not more than 2 mm (0.08 in.) in the empty incubating chamber with shelves in place, and vents open. Locate one thermocouple in each of the eight corners of the incubator approximately 5 cm (2 in.) from each wall, and place the ninth thermocouple within 2.5 cm (1 in.) of the geometric center of the chamber. A minimum length of 30 cm (12 in.) of wire for each thermocouple shall be inside the incubator to minimize the conduction of heat from the thermocouple.

NOTE 1—If calibrated thermocouples are not available, nine thermocouples made from the same spool of wire may be used, provided they give the same value for temperature when placed adjacent to one another in the incubating chamber at the temperature of the test.

4.1.2 Bring the incubator to the specified temperature and allow it to reach a steady state (Note 2). Record the temperature of the nine thermocouples for a period of at least 12 h and determine from the record the maximum deviation of each point from the desired temperature. The ambient room temperature shall vary by not more than a total of 5 °C and the line voltage shall not vary by more than a total of 5 % during the test.

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.06 on Laboratory Instruments and Equipment.

Current edition approved Jan. 1, 2022. Published February 2022. Originally approved in 1989. Last previous edition approved in 2016 as E1292 – 94 (2016). DOI: 10.1520/E1292-94R22.