



Designation: E1584 – 23

Standard Test Method for Assay of Nitric Acid¹

This standard is issued under the fixed designation E1584; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers determination of the assay of nitric acid by total acidity.

1.2 This test method is suitable for concentrations between approximately 50 % and 70 %, calculated as nitric acid.

1.3 The following applies for the purposes of determining the conformance of the test results using this test method to applicable specifications, results shall be rounded off in accordance with the rounding-off method of Practice E29.

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 8.

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

D1193 Specification for Reagent Water

D6809 Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic, Industrial, Specialty and Related Chemicals and is the direct responsibility of Subcommittee D16.10 on Acids.

Current edition approved July 1, 2023. Published July 2023. Originally approved in 1993. Last previous edition approved in 2017 as E1584 – 17. DOI: 10.1520/E1584-23.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

E180 Practice for Determining the Precision of ASTM Methods for Analysis and Testing of Industrial and Specialty Chemicals (Withdrawn 2009)³

E200 Practice for Preparation, Standardization, and Storage of Standard and Reagent Solutions for Chemical Analysis

E300 Practice for Sampling Industrial Chemicals

2.2 *Other Documents:*

OSHA Regulations, 29 CFR, paragraphs 1910.1000 and 1910.1200⁴

3. Summary of Test Method

3.1 A weighed sample of acid is diluted in water and titrated with 1.0 N sodium hydroxide solution, using phenolphthalein as the end-point indicator.

4. Significance and Use

4.1 This test method provides a means for assaying nitric acid, based on total acidity. The concentration of nitric acid is important in many of the uses of nitric acid, including specification compliance and manufacturing control.

5. Interferences

5.1 Acids other than nitric, and compounds that consume sodium hydroxide, will yield erroneously high results.

6. Apparatus

6.1 *Erlenmeyer Flask*, 250 mL, glass stoppered.

6.2 *Buret*, 50 mL, Class A.

NOTE 1—A digital buret or automated titrator capable of measuring volumes to the nearest 0.01 mL may be used in place of a conventional buret.

7. Reagents

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where

³ The last approved version of this historical standard is referenced on www.astm.org.

⁴ Available from Superintendent of Documents, U.S. Government Printing Office, 732 N. Capitol St. NW, Mail Stop: SDE, Washington, DC 20401.

*A Summary of Changes section appears at the end of this standard