



Standard Practice for Personal Sampling and Analysis of Endotoxin in Metalworking Fluid Aerosols in Workplace Atmospheres¹

This standard is issued under the fixed designation E2144; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ε) indicates an editorial change since the last revision or reappraisal.

1. Scope*

1.1 This practice covers quantitative methods for the personal sampling and determination of bacterial endotoxin concentrations in poly-disperse metalworking fluid aerosols in workplace atmospheres. Users should have fundamental knowledge of microbiological techniques and endotoxin testing.

1.2 Users of this practice may obtain personal or area exposure data of endotoxin in metalworking fluid aerosols, either on a short-term or full-shift basis in workplace atmospheres.

1.3 This practice gives an estimate of the endotoxin concentration of the sampled atmosphere.

1.4 This practice seeks to minimize interlaboratory variation but does not ensure uniformity of results.

1.5 It is anticipated that this practice will facilitate interlaboratory comparisons of airborne endotoxin data from metalworking fluid atmospheres, particularly metal removal fluid atmospheres, by providing a basis for endotoxin sampling, extraction, and analytical methods.

1.6 In 1997, the Occupational Safety and Health Administration (OSHA) empanelled a Standards Advisory Committee to make recommendations to the Administration regarding measures that the Administration could take to improve the health of workers exposed to metalworking fluids. A report to the Assistant Secretary of Labor for OSHA was submitted in July 1999. Subcommittee E34.50 believes that the user community would benefit significantly if a standard method was developed to give the community guidance on a methodology for the sampling and analysis of personal airborne endotoxin exposure assessments in facilities using water-miscible metal removal fluids, based on the LAL assay or other endotoxin detection technologies as they become available.

¹ This practice is under the jurisdiction of ASTM Committee E34 on Occupational Health and Safety and is the direct responsibility of Subcommittee E34.50 on Health and Safety Standards for Metal Working Fluids.

Current edition approved Nov. 1, 2021. Published December 2021. Originally approved in 2001. Last previous edition approved in 2016 as E2144 – 11 (2016). DOI: 10.1520/E2144-21.

1.7 This practice does not attempt to set or imply limits for personal exposure to endotoxin in metalworking fluid aerosols in workplace environments.

1.8 *Units*—The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.9 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.10 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- D1356 Terminology Relating to Sampling and Analysis of Atmospheres
- D2881 Classification for Metalworking Fluids and Related Materials
- D4840 Guide for Sample Chain-of-Custody Procedures
- D5337 Practice for Flow Rate Adjustment of Personal Sampling Pumps
- E1370 Guide for Air Sampling Strategies for Worker and Workplace Protection
- E1497 Practice for Selection and Safe Use of Water-Miscible and Straight Oil Metal Removal Fluids
- E1542 Terminology Relating to Occupational Health and Safety
- E2523 Terminology for Metalworking Fluids and Operations

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

2.2 Government Standards:³

29 CFR 1910.1000 Air Contaminants

29 CFR 1910.1450 Occupational Exposure to Hazardous Chemicals in Laboratories

2.3 Other Documents:

Criteria Document for a Recommended Standard: Occupational Exposure to Metalworking Fluids⁴

NIOSH Manual of Analytical Methods (NMAM), 5th Edition⁵

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms relating to this practice, refer to Terminologies D1356, E1542, and E2523.

3.1.2 *control standard endotoxin (CSE)*, *n*—a purified preparation of endotoxin based on the USP Reference Standard Endotoxin (RSE); used in laboratories to prepare standard solutions.

3.1.3 *endotoxin*, *n*—pyrogenic high molar mass lipopolysaccharide (LPS) complex associated with the cell wall of Gram-negative bacteria.

3.1.3.1 *Discussion*—Though endotoxins are pyrogens, not all pyrogens are endotoxins. Endotoxins are specifically detected through a Limulus Amoebocyte Lysate (LAL) test.

3.1.4 *endotoxin unit (EU)*, *n*—a biological potency unit equivalent to the FDA Reference Standard Endotoxin (RSE).

3.1.4.1 *Discussion*—The current RSE (EC-6) is equivalent to 1 ng = 10 EU.

3.1.5 *Gram-negative bacteria*, *n*—prokaryotic cells that have a complex cell wall structure that stains characteristically when subjected to the differential Gram staining procedure.

3.1.6 *Limulus amoebocyte lysate (LAL) assay*, *n*—a biological assay dependent on a series of cascading enzyme reactions that occur when Limulus blood cell (amebocyte) lysate combines with endotoxin.

3.1.7 *metal removal fluid (MRF)*, *n*—any fluid in the subclass of metalworking fluids used to cut or otherwise take away material or piece of stock.

3.1.7.1 *Discussion*—Metal removal fluids include straight or neat oils (Classification D2881) not intended for further dilution with water, and water-miscible soluble oils, semisynthetics, and synthetics, which are intended to be diluted with water before use.

3.1.7.2 *Discussion*—Metal removal fluids become contaminated during use in the workplace with a variety of workplace substances including, but not limited to: abrasive particles, tramp oils, cleaners, dirt, metal fines and shavings, dissolved

metal and hard water salts, bacteria, fungi, microbiological decay products, and waste. These contaminants can cause changes in the lubricity and cooling ability of the metal removal fluid as well as have the potential to adversely affect the health and welfare of employees in with the contaminated metal removal fluid.

3.1.8 *personal sampler*, *n*—a portable sampling instrument that is attached to a person to ascertain the concentration of specific constituents in the air in the person's breathing zone.

3.1.9 *pyrogen-free*, *adj*—material(s) devoid of measurable endotoxin activity.

3.1.10 *pyrogen-free water (PFW)*, *n*—processed water that is devoid of measurable endotoxin activity.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *field blank*, *n*—filter/cassette unit prepared for sampling that is taken to the sampling site and handled in the same manner as the analytical filter/cassette unit, but that is not a part of the sampling process.

3.2.2 *onset time*, *n*—time required for a change of 200 mOD (optical density) units relative to the initial OD value.

4. Summary of Practice

4.1 A known volume of workplace air in a facility utilizing metalworking fluids is drawn through a sample filter cassette unit.

4.2 The sample filter is extracted into a pyrogen-free solution to quantitatively release endotoxin absorbed from collected metalworking fluid aerosol.

4.3 The extract solution is subjected to quantitative endotoxin analysis techniques. The measured endotoxin concentration is reported in terms of endotoxin potency units per unit volume of air sampled.

5. Significance and Use

5.1 Endotoxins in metalworking fluid aerosols present potential respiratory health hazards to workers who inhale them. Therefore, a consensus standard is needed to provide reliable data on workplace airborne endotoxin concentrations where metalworking fluids are used.

5.2 This practice for measuring airborne endotoxin concentrations in metalworking fluid atmospheres will help to foster a better understanding of endotoxin exposure-response relationships.

5.3 This practice facilitates comparisons of interlaboratory data from methods and field investigative studies.

6. Interferences

6.1 Airborne endotoxin measurements resulting from use of LAL reagents are subject to inhibition/enhancement effects from a variety of biomolecular species and physicochemical phenomena, such as pH, temperature, filter matrix effects, cationic concentrations, LAL-reactive materials (LRM), enzyme influences, and lysate composition variability and sensitivity (a function of different lysate processing methodologies).

³ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.

⁴ Available from U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, 4676 Columbia Pkwy., Cincinnati, OH 45226, <https://www.cdc.gov/niosh/docs/98-102/default.html>.

⁵ Available from U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, 4676 Columbia Pkwy., Cincinnati, OH 45226, <https://www.cdc.gov/niosh/nmam/default.html>.