



Standard Test Method for Calibration of Thermocouples By Comparison Techniques¹

This standard is issued under the fixed designation E220; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope

1.1 This test method describes the principles, apparatus, and procedure for calibrating thermocouples by comparison with a reference thermometer. Calibrations are covered over temperature ranges appropriate to the individual types of thermocouples within an overall range from approximately –195 to 1700 °C (–320 to 3100 °F).

1.2 In general, this test method is applicable to unused thermocouples. This test method does not apply to used thermocouples due to their potential material inhomogeneity—the effects of which cannot be identified or quantified by standard calibration techniques. Thermocouples with large-diameter thermoelements and sheathed thermocouples may require special care to control thermal conduction losses.

1.3 In this test method, all values of temperature are based on the International Temperature Scale of 1990. See Guide E1594.

1.4 *This standard may involve hazardous materials, operations and equipment. This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory requirements prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

- E1 Specification for ASTM Liquid-in-Glass Thermometers
- E77 Test Method for Inspection and Verification of Thermometers
- E230 Specification and Temperature-Electromotive Force (EMF) Tables for Standardized Thermocouples

¹ This test method is under the jurisdiction of ASTM Committee E20 on Temperature Measurement and is the direct responsibility of Subcommittee E20.04 on Thermocouples.

Current edition approved Nov. 1, 2013. Published December 2013. Originally approved in 1963. Last previous edition approved in 2007 as E220 – 07A. DOI: 10.1520/E0220-13.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- E344 Terminology Relating to Thermometry and Hydrometry
- E452 Test Method for Calibration of Refractory Metal Thermocouples Using a Radiation Thermometer
- E563 Practice for Preparation and Use of an Ice-Point Bath as a Reference Temperature
- E644 Test Methods for Testing Industrial Resistance Thermometers
- E1129/E1129M Specification for Thermocouple Connectors
- E1594 Guide for Expression of Temperature
- E1684 Specification for Miniature Thermocouple Connectors
- E1751 Guide for Temperature Electromotive Force (EMF) Tables for Non-Letter Designated Thermocouple Combinations (Withdrawn 2009)³
- E2846 Guide for Thermocouple Verification

3. Terminology

3.1 *Definitions*—The definitions given in Terminology E344 shall apply to this test method.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *check standard, n*—a measurement instrument or standard whose repeated results of measurement are used to determine the repeatability of a calibration process and to verify that the results of a calibration processes are statistically consistent with past results.

3.2.2 *isothermal block, n*—a piece of solid material of high thermal conductivity used to promote thermal equilibrium between two or more thermometers.

3.2.3 *reference junction compensation, n*—the adjustment of the indication of a thermocouple such that the adjusted indication is equivalent to the emf or temperature that the thermocouple would indicate if the reference junctions were maintained at 0°C.

3.2.3.1 *Discussion*—In most cases, the thermocouple indication is adjusted by measuring the temperature of a terminal block where the thermocouple is connected, and then adding to the thermocouple emf an additional emf equal to the emf of the

³ The last approved version of this historical standard is referenced on www.astm.org.

thermocouple reference function evaluated at the temperature of the terminal block. Because the emf-temperature relationship of any actual thermocouple differs slightly from that of the reference function, reference junction compensation typically introduces higher uncertainties compared to the use of a well-prepared ice bath.

3.2.4 *reference junction compensator, n*—a device that implements reference junction compensation.

3.2.5 *reference thermometer, n*—thermometer that establishes the value of temperature in a given system containing additional temperature sensors.

3.2.5.1 *Discussion*—In a calibration system the reference thermometer is a calibrated thermometer capable of indicating values of temperature with known uncertainty. The reference thermometer provides the standard temperature for the system at the time of test.

3.2.6 *thermocouple type, n*—a standardized thermoelectric class of thermoelement materials that, used as a pair, have a normal relationship between relative Seebeck emf and temperature.

3.2.6.1 *Discussion*—For common, commercially available thermocouples, a thermocouple type is identified by a letter designation (types B, C, E, J, K, N, R, S, and T). The letter designation scheme is given in Guide E2846. The tables in E1751 and E1751 give temperature-EMF relationships for a number of additional thermocouple compositions that are not identified by a letter designation.

4. Summary of Test Method

4.1 Comparison calibration consists of measuring the emf of the thermocouple being calibrated in an isothermal medium while simultaneously measuring the temperature of the medium with a reference thermometer. The reference thermometer may be any thermometer with sufficient accuracy at the temperature of calibration.

5. Significance and Use

5.1 For users or manufacturers of thermocouples, this test method provides a means of verifying the emf-temperature characteristics of the material prior to use.

5.2 This test method can be used to calibrate a thermocouple for use as a reference, or it can be used to calibrate thermocouples representing a batch of purchased, assembled thermocouples.

5.3 This test method can be used for the verification of the conformance of thermocouple materials to temperature tolerances for specifications such as the tables in Specification E230 or other special specifications as required for commercial, military, or research applications.

6. Interferences

6.1 Since the success of this test method depends largely upon the ability to maintain the measuring junction of the thermocouple being calibrated and the reference thermometer at the same temperature, considerable care must be taken in choosing the media and conditions under which the comparisons are made. Stirred liquid baths, uniformly heated metal

blocks, tube furnaces, and dry fluidized baths, properly used, are acceptable temperature comparison environments. In the case of large diameter thermoelements and sheathed thermocouples, special attention must be given to effects of thermal conduction.

6.2 Voltage measurement instruments with sufficiently high input impedance must be used for measuring thermocouple emf to eliminate instrument loading as a significant source of error. The ratio of input impedance to thermocouple loop resistance should be significantly (at least 10^4) greater than the ratio of the measured emf to the desired emf uncertainty.

6.3 The test method relies on the assumption that test thermoelements are homogeneous. If so, their output voltage at a given measuring junction temperature is independent of temperature variations along the length of the thermocouple. Departures from this ideal contribute to uncertainty in the use of test results. The effects typically are negligibly small for new, unused thermocouple material, but not for used thermocouples, especially those of base-metal composition. The effects of inhomogeneity can be identified, but not accurately quantified, by the techniques described in Appendix X4 in this test method and in section 8.2 of Guide E2846. Descriptions of the testing of used thermocouples may be found Guide E2846 and Manual MNL 12 (1).⁴

6.4 This test method presumes that the tested thermocouples are suitable for use in air throughout the range of calibration temperatures. To avoid oxidation of the thermoelements, refractory-metal thermocouples that have not been hermetically sealed in a sheath suitable for use in air should be tested in an inert gas environment at temperatures above approximately 500 °C. In this case, use of this test method is recommended in combination with the furnaces and related procedures described in Test Method E452.

7. Apparatus

7.1 The choice of apparatus used for the comparison test will depend primarily on the temperature range to be covered and on the desired calibration uncertainty. The apparatus required for the application of this test method will depend in detail upon the temperature range being covered but in all cases shall be selected from the equipment described as follows:

7.2 *Comparison Baths and Furnaces*—A controlled temperature comparison medium (bath or furnace) shall be used in which the measuring junction of the thermocouple to be calibrated is brought to the same temperature as a reference thermometer. The spatial uniformity of temperature within the nominally isothermal calibration zone shall be established. Acceptable methods include measurements of the calibration zone at the time of testing or the use of control charts that display the periodic calibration of check standards or the periodic characterization of the calibration zone. The frequency of such testing will depend on the inherent stability of the bath or furnace. The uniformity of the calibration zone shall be remeasured sufficiently often such that any deviations in

⁴ The boldface numbers in parenthesis refer to the list of references at the end of this standard.