



Standard Test Method for Measurement of Adenosine Triphosphate in Water-Miscible Metalworking Fluids¹

This standard is issued under the fixed designation E2694; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ε) indicates an editorial change since the last revision or reappraisal.

1. Scope*

1.1 This test method provides a protocol for capturing, extracting, and quantifying the adenosine triphosphate (ATP) content associated with microorganisms found in water-miscible metalworking fluids (MWFs).

1.2 The ATP is measured using a bioluminescence enzyme assay, whereby light is generated in amounts proportional to the concentration of ATP in the samples. The light is produced and measured quantitatively as relative light units (RLUs) which are converted by comparison with an ATP standard and computation to pg ATP/mL.

1.3 This test method is equally suitable for use in the laboratory or field.

1.4 The test method detects ATP concentrations in the range of 4.0 pg ATP/mL to 400 000 pg ATP/mL.

1.5 Providing interferences can be overcome, bioluminescence is a reliable and proven method for qualifying and quantifying ATP. The method does not differentiate between ATP from different sources, for example, from different types of microorganisms, such as bacteria and fungi.

1.6 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.8 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee E34 on Occupational Health and Safety and is the direct responsibility of Subcommittee E34.50 on Health and Safety Standards for Metal Working Fluids.

Current edition approved Nov. 1, 2021. Published November 2021. Originally approved in 2009. Last previous edition approved in 2016 as E2694 – 16. DOI:10.1520/E2694-21.

2. Referenced Documents

2.1 ASTM Standards:²

- D1129 Terminology Relating to Water
 - D4012 Test Method for Adenosine Triphosphate (ATP) Content of Microorganisms in Water
 - D4840 Guide for Sample Chain-of-Custody Procedures
 - D6161 Terminology Used for Microfiltration, Ultrafiltration, Nanofiltration, and Reverse Osmosis Membrane Processes
 - E177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods
 - E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method
 - E1326 Guide for Evaluating Non-culture Microbiological Tests
 - E1497 Practice for Selection and Safe Use of Water-Miscible and Straight Oil Metal Removal Fluids
 - E2523 Terminology for Metalworking Fluids and Operations
- ### 2.2 Government Standards:³
- 29 CFR 1910.1000 Air Contaminants
 - 29 CFR 1910.1450 Occupational Exposure to Hazardous Chemicals in Laboratories

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this test method, refer to Terminologies D1129, D6161, and E2523.

3.1.2 *adenosine monophosphate (AMP), n*—the molecule formed by the removal of two molecules of phosphate (one pyrophosphate molecule) from ATP.

3.1.3 *adenosine triphosphate (ATP), n*—a molecule comprised of a purine and three phosphate groups that serves as the primary energy transport molecule in all biological cells.

3.1.4 *aseptic, adj*—sterile, free from viable microbial contamination.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.

*A Summary of Changes section appears at the end of this standard

3.1.5 *bioluminescence*, *n*—the production and emission of light by a living organism as the result of a chemical reaction during which chemical energy is converted to light energy.

3.1.6 *biomass*, *n*—any matter which is or was a living organism or excreted from a microorganism (D6161).

3.1.7 *culturable*, *adj*—microorganisms that proliferate as indicated by the formation of colonies on solid growth media or the development of turbidity in liquid growth media under specific growth conditions.

3.1.8 *Luciferase*, *n*—a general term for a class of enzymes that catalyze bioluminescent reactions.

3.1.9 *Luciferin*, *n*—a general term for a class of light-emitting biological pigments found in organisms capable of bioluminescence.

3.1.10 *luminometer*, *n*—an instrument capable of measuring light emitted as a result of non-thermal excitation.

3.1.11 *relative light unit (RLU)*, *n*—an instrument-specific unit of measurement reflecting the number of photons emitted by the Luciferin-Luciferase driven hydrolysis of ATP to AMP plus pyrophosphate.

3.1.11.1 *Discussion*—RLU is not an SI unit, however, RLU is proportional to ATP concentration.

3.1.12 *viable microbial biomass*, *n*—metabolically active (living) microorganisms.

3.2 Acronyms:

3.2.1 *AMP*—adenosine monophosphate

3.2.2 *ATP*—adenosine triphosphate

3.2.3 *HDPE*—high density polyethylene

3.2.4 *MWF*—metalworking fluid

3.2.5 *PP*—polypropylene

3.2.6 *RLU*—relative light unit

4. Summary of Test Method

4.1 A control assay is performed using 100 μL of 1.0 ng ATP/mL standard.

4.2 A 5.0 mL sample of MWF is placed into a syringe and then pressure filtered through a 0.7 μm , glass-fiber, in-line depth filter.

4.3 The retentate is then washed with a reagent to remove extracellular ATP and other contaminants that might otherwise interfere with the ATP assay.

4.4 The filter is air dried.

4.5 A lysing reagent is used to release ATP from microbial cells that have been captured on the glass-fiber filter, and the filtrate is dispensed into an unused culture tube.

4.6 The filtrate is diluted 1+9 with a buffer solution.

4.7 A 100 μL volume of diluted filtrate is transferred to an unused culture tube into which 100 μL of Luciferin-Luciferase reagent has previously been dispensed.

4.8 The culture tube is placed into a luminometer and the light intensity is read in RLU.

4.9 RLU is converted to Log_{10} [pg ATP/mL] of sample by computation.

4.10 A procedure for differentiating between bacterial and fungal cATP biomass is provided in [Appendix X4](#).

4.11 A procedure for determining the total ATP (tATP) biomass on MWF system surfaces is provided in [Appendix X5](#).

5. Significance and Use

5.1 This method measures the concentration of ATP present in the sample. ATP is a constituent of all living cells, including bacteria and fungi. Consequently, the presence of ATP is an indicator of total microbial contamination in metalworking fluids. ATP is not associated with matter of non-biological origin.

5.2 Test Method [D4012](#) validated ATP as a surrogate for culturable bacterial data (Guide [E1326](#)).

5.3 This method differs from Test Method [D4012](#) in that it eliminates interferences that have historically rendered ATP testing unusable with complex organic fluids such as MWFs.

5.4 The ATP test provides rapid test results that reflect the total bioburden in the sample. It thereby reduces the delay between test initiation and data capture, from the 36 h to 48 h (or longer) required for culturable colonies to become visible, to approximately 5 min.

5.5 Although ATP data generally covary with culture data in MWF,⁴ different factors affect ATP concentration than those that affect culturability.

5.5.1 Culturability is affected primarily by the ability of captured microbes to proliferate on the growth medium provided, under specific growth conditions. It has been estimated that less than 1 % of the species present in an environmental sample will form colonies under any given set of growth conditions.⁵

5.5.2 ATP concentration is affected by: the microbial species present, the physiological states of those species, and the total bioburden (see [Appendix X1](#)).

5.5.2.1 One example of the species effect is that the amount of ATP per cell is substantially greater for fungi than bacteria.

5.5.2.2 Within a species, cells that are more metabolically active will have more ATP per cell than dormant cells.

5.5.2.3 The greater the total bioburden, the greater the ATP concentration in a sample.

5.5.3 The possibility exists that the rinse step ([11.15](#)) may not eliminate all chemical substances that can interfere with the bioluminescence reaction ([11.39](#)).

5.5.3.1 The presence of any such interferences can be evaluated by performing a standard addition test series as described in [Appendix X3](#).

5.5.3.2 Any impact of interfering chemicals will be reflected as bias relative to data obtained from fluid that does not contain interfering chemicals.

⁴ Passman, et al., "Real-Time Testing of Bioburdens in Metalworking Fluids using Adenosine Triphosphate as a Biomass Indicator," 2009 STLE Annual Meeting, Orlando, FL.

⁵ Sloan, W. T., Quince, C., and Curtis, T. P., "The Uncountables," *Accessing Uncultivated Microorganisms*, ASM Press, Washington, DC, 2008, p. 35.