



Designation: E769 – 22

Standard Test Methods for Odor of Methanol, Ethanol, *n*-Propanol, and Isopropanol¹

This standard is issued under the fixed designation E769; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 These test methods cover procedures for observing characteristic and residual odors and estimating odor intensity of methanol, ethanol, *n*-propanol, and isopropanol for the purpose of determining their odor acceptability. Several test methods are included and any or all may be used. These test methods may be used for both denatured and undenatured alcohols.

1.2 These test methods are not intended to determine subtle odor differences between alcohols, which do not affect their end use.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

D86 Test Method for Distillation of Petroleum Products and Liquid Fuels at Atmospheric Pressure
E133 Specification for Distillation Equipment

3. Significance and Use

3.1 These test methods are intended to provide procedures that may be used to evaluate denatured and undenatured alcohols with 1 to 3 carbon atoms for characteristic and residual odors that will affect their end use.

¹ These test methods are under the jurisdiction of ASTM Committee E18 on Sensory Evaluation and are the direct responsibility of Subcommittee E18.06 on Food and Beverage Evaluation.

Current edition approved March 1, 2022. Published March 2022. Originally approved in 1981. Last previous edition approved in 2004 as E769 – 85 (2004) which was withdrawn August 2011 and reinstated in March 2022.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3.2 The products tested are compared to reference standards using one or all of the test methods described. Interpretation of results and acceptability of the product for a specific end use are the responsibility of the panel supervisor and are based on the results obtained in testing.

3.3 These tests may be used to qualify the suitability of these products for use in a customer-producer relationship.

4. Reference Standards

4.1 Reference standards of the products to be tested, having odor characteristics satisfactory to the consumer and producer, are required.

4.2 Periodically examine reference standards to assure their continued acceptability, and replace when necessary.

5. Selection of Panel

5.1 A panel of at least five trained judges is recommended for these evaluations.

5.2 The panel may be selected and trained in accordance with the methods outlined in ASTM STP 758.³ Since discrimination tests are involved, panel members must have a complete understanding of the nature of the judgements required, the test procedures, and test controls required. During the training period, panelists should examine both acceptable and unacceptable products to become familiar with the range of acceptability.

6. Procedure A—Characteristic and Residual Odors by Evaporation

6.1 Apparatus:

6.1.1 *Rapid Qualitative Filter Paper*, 2.5 cm by 7.5 cm strips, odor-free.

6.1.2 *Glass Containers*, odor-free, with fitted closures.

6.2 *Sample Preparation*—Each panelist dips separate strips of the filter paper to a depth of 5.0 cm into the sample and standard, which are provided in suitable, clean, odor-free glass containers with fitted closures.

6.3 Sample Evaluation:

³ *Selection and Training of Sensory Evaluation Panels. ASTM STP 758*, ASTM International, 1981.