

IEC 60789
(Third edition – 2005)

**Medical electrical equipment –
Characteristics and test conditions of radionuclide imaging devices –
Anger type gamma cameras**

CORRIGENDUM 1

Table 1 – Radionuclides and energy windows to be used for performance measurements

Instead of:

**Table 1 – Radionuclides and energy windows
to be used for performance measurements**

Radionuclide	Energy window keV
^{99m}Tc	141, with a tolerance of $\pm 7,5 \%$
^{131}I	364, with a tolerance of $\pm 10 \%$
^{67}Ga	93, 184, 300, with a tolerance of $\pm 20 \%$
^{68}Ga	511, with a tolerance of $\pm 10 \%$
NOTE Because the characteristics of a gamma camera may change noticeably between 122 keV (^{57}Co) and 141 keV (^{99m}Tc), the former is not included as a suitable radionuclide. However, it may be useful in some circumstances, e.g. for quality control.	

read:

**Table 1 – Radionuclides and energy windows
to be used for performance measurements**

Radionuclide	Energy window keV
^{99m}Tc	141 ($\pm 7,5 \%$)
^{131}I	364 ($\pm 10 \%$)
^{67}Ga	93, 184, 300 ($\pm 20 \%$)
^{68}Ga	511 ($\pm 10 \%$)
NOTE Because the characteristics of a gamma camera may change noticeably between 122 keV (^{57}Co) and 141 keV (^{99m}Tc), the former is not included as a suitable radionuclide. However, it may be useful in some circumstances, e.g. for quality control.	