



Optics and optical instruments — Visual acuity testing — Method of correlating optotypes

TECHNICAL CORRIGENDUM 1

Optique et instruments d'optique — Méthode d'essai de l'acuité visuelle — Méthode de corrélation entre les optotypes

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 8597:1994 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic, endoscopic, metrological instruments and test methods*.

Page 1

Subclause 4.2 shall read:

"4.2 Test area

The test area shall be circular with a diameter of $4^\circ \pm 0,4^\circ$.

The surrounding field shall have a diameter of $15^\circ \pm 1,5^\circ$ and shall be illuminated homogeneously so that it does not influence the measurement.

The luminance of the surrounding field shall not be greater than that of the test area."