



INTERNATIONAL STANDARD ISO 8761:1989
TECHNICAL CORRIGENDUM 1

Published 2009-07-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Work-place air — Determination of mass concentration of
nitrogen dioxide — Method using detector tubes for short-term
sampling with direct indication**

TECHNICAL CORRIGENDUM 1

*Air des lieux de travail — Détermination de la concentration en masse du dioxyde d'azote — Méthode utilisant
des tubes détecteurs pour échantillonnage rapide à lecture directe*

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 8761:1989 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

Page 4, 9.2, Paragraph 1

Insert a new third sentence, after "... 6 mg/m³",

"The reading error must be lower than 4 ppm (4 ml per 10⁶ ml), otherwise the maximum random error will be bigger than 25 % (expressed as two times the coefficient of variation)."