



INTERNATIONAL STANDARD ISO 9339-2:1998 TECHNICAL CORRIGENDUM 1

Published 2000-04-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Optics and optical instruments — Contact lenses — Determination of thickness

Part 2: Hydrogel contact lenses

TECHNICAL CORRIGENDUM 1

Optique et instruments d'optique — Lentilles de contact — Détermination de l'épaisseur — Partie 2: Lentilles de contact en hydrogel

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 9339-2:1998 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

Page 3

Clause 6 Conditioning of contact lenses before testing

In the second line, replace "...at a temperature of 20,0 °C ± 5 °C..." with "...at a temperature of 20,0 °C ± 0,5 °C...".